

Recessed AED Cabinet with Alarm

Instructions for use · HS-CAB-200

Before use

Check the readiness indicator, the expiry date on the pack and that the item matches the device it is intended for. Do not use an item that is past its expiry date or whose packaging is damaged.

Use

Follow the voice prompts of the device. Replace consumables after any use, and record the replacement in your AED programme so the next readiness report is correct.

Storage

Store between 32 and 109 °F (0 to 43 °C), out of direct sunlight. Keep the item in its sealed packaging until it is needed.

Disposal

Dispose of used items in line with local regulations for medical consumables.

Demo document. HeartStream demo environment, generated 2026-09-22 for a Truvio product demonstration. Not a real instruction for use and not for clinical reference.